# Smart Probe Calibration Guide

Is the calibration for your VFC 800-WiFi data logger smart probe expiring soon? Here are the steps to purchase a newly calibrated smart probe.

**Ordering replacement smart probe:**

1. Visit https://www.vfcdataloggers.com/product/vfc800wifi-replacement-probe/ 1-2 months prior to calibration expiration date to allow time for shipment and installation of the new probe.
2. Choose the 10ft probe cord length option. Enter quantity needed, add to cart and check out. Add shipping information and payment method.
   - ❖ *New probes cannot be shipped to PO Box, must be to a physical address. Please allow up to 2 weeks for delivery.*




- ➢ **No Recalibration Downtime:** A major advantage of smart probe data loggers is their seamless recalibration process. When calibration expires, probes can be instantly swapped without interrupting operations or sending the logger for recalibration, ensuring uninterrupted monitoring and data collection.


Once the calibrated probe is received, follow the directions below for installation:

1. Unplug the datalogger and set aside.
2. Remove buffered solution bottle from fridge/freezer, remove cap and separate the expired probe from the cap.
3. Plug new probe into your datalogger.
4. Insert steel probe into the cap of the existing buffered solution ensuring the plug is the correct direction and replace cap with inserted probe to the buffered solution bottle.
5. Label your new calibration certificates with the name of the unit (fridge/freezer) they are in and keep them available for your site visits.









- ➢ For a video tutorial on how to install your new probe visit: https://www.youtube.com/watch?v=2Vxpj6wytUc&t=3s


**Maine Immunization Program**
**Tel: (207) 207-287-3746; Fax: (207) 287-8127**
**Email:** ImmunizeME.DHHS@maine.gov